



# drexel® electronics & engineering products (p) ltd

Works:442 ,JLPL,Industrial Area  
Sector 82, Mohali (PB) 160055  
Ph: +91172 4190098  
Mob: +91 9988610098  
Email: sales@drexel.in  
Web: www.drexel.in

Corp: 13/2, 2<sup>nd</sup>& 3<sup>rd</sup>Floor  
Opp. Bus terminal Mehrauli,  
New Delhi 110030  
Toll Free:1800 123440098  
Mob: +91 9316660098  
info@drexel.in

## Calibration Certificate

<b>Sample Code</b>	DXL/2024/30506	<i>Page No.: 01/01</i>	
<b>Customer Name &amp; Address:</b>	M/s Haat Supreme Wastech Private Limited #V.P.O. Bazida Jatan, Karnal, Haryana-132037		
<b>Calibration Date:</b>	20-06-2024	<b>Calibration Due Date:</b>	19-12-2024

### PRODUCT DETAIL:


<b>Product Name:</b>	Online Emission Monitoring System	<b>Make &amp; Model:</b>	Drexel & DX4B
<b>Product Sr. No.:</b>	53	<b>Voltage &amp; Frequency:</b>	220V AC/50 Hz

### Results:

Sr. No.	Sensor Applicable	Gas Conc.	Current (Range 4 mA - 20 mA )			Remark
			Minimum	Span	Maximum	
1	CO	500 PPM	0 PPM	502 PPM	10,000 PPM	ok
2	CO2	6 %	0 PPM	5.3 %	12 %	ok
3	HCL	500 PPM	0 PPM	498 PPM	1000 PPM	ok
4	NOX	500 PPM	0 PPM	495 PPM	1000 PPM	ok
5	SPM	Zero Check		Isokinetic Technique		
<b>Suction Unit:</b>						
1.	Suction Pump	ok	Auto- clean Filter Assembly			ok
2.	Filter Assembly	ok	Purging Cycle			ok

**Remarks:** System is completely tested & inspected properly and found OK.

  
20/06/2024  
Tested By  
Mr. Israfil  
Field Engineer

  
Checked By  
Mr. Gurpreet Singh Field  
Senior Engineer (Technical)

