



drexel® electronics & engineering products (p) ltd

Works:442 ,JLPL,Industrial Area
Sector 82, Mohali (PB) 160055
Ph: +91172 4190098
Mob: +91 9988610098
Email: sales@drexel.in
Web: www.drexel.in

Corp: 13/2, 2nd& 3rdFloor
Opp. Bus terminal Mehrauli,
New Delhi 110030
Toll Free:1800 123440098
Mob: +91 9316660098
info@drexel.in

Calibration Certificate

| | | | |
|-------------------------------------|--|------------------------------|------------|
| Sample Code | DXL/2025/1108 | <i>Page No.: 01/01</i> | |
| Customer Name & Address: | M/s Haat Supreme Wastech Private Limited #V.P.O. Bazida Jatan, Karnal, Haryana-132037 | | |
| Calibration Date: | 26-03-2025 | Calibration Due Date: | 26-09-2025 |

PRODUCT DETAIL:

| | | | |
|-------------------------|-----------------------------------|---------------------------------|---------------|
| Product Name: | Online Emission Monitoring System | Make & Model: | Drexel & DX4B |
| Product Sr. No.: | 53 | Voltage & Frequency: | 220V AC/50 Hz |

Results:

| Sr. No. | Sensor Applicable | Gas Conc. | Current (Range 4 mA - 20 mA) | | | Remark |
|----------------------|-------------------|------------|-------------------------------|----------------------|------------|--------|
| | | | Minimum | Span | Maximum | |
| 1 | CO | 500 PPM | 0 PPM | 503 PPM | 10,000 PPM | ok |
| 2 | CO2 | 6 % | 0 PPM | 5.1 % | 12 % | ok |
| 3 | HCL | 500 PPM | 0 PPM | 500 PPM | 1000 PPM | ok |
| 4 | NOX | 500 PPM | 0 PPM | 495 PPM | 1000 PPM | ok |
| 5 | SPM | Zero Check | | Isokinetic Technique | | |
| Suction Unit: | | | | | | |
| 1. | Suction Pump | ok | Auto- clean Filter Assembly | | | ok |
| 2. | Filter Assembly | ok | Purging Cycle | | | ok |

Remarks: System is completely tested & inspected properly and found OK.

Tested By
Mr. Israfil
Field Engineer



Checked By
Mr. Gurpreet Singh Field
Senior Engineer (Technical)

